

EC Declaration of Product Conformity

The manufacturer established in the community;

**BM Polyco Ltd
Crown Road
Enfield
EN1 1TX
United Kingdom**

Certified to ISO 9001 & ISO 13485

declares that the new medical device described hereafter



VITRILE PF GL663

Non-sterile, powder-free, nitrile/vinyl disposable glove

is in conformity with the provisions of Council Directive 93/42/EEC and with National Standards transposing harmonised standards EN455-1, EN455-2, EN455-3 and EN455-4

is Class I and is self-certified according to the requirements of EC Directive 93/42/EEC and subsequent amendments

Product Information

Glove Care: Store below 40°C away from direct sunlight

Freedom from holes: AQL 1.5

Physical properties: 3.6N minimum



User Information

This product does not contain natural rubber latex.

This product contains low levels of residual chemical accelerators, which may cause allergic reactions.

Range Available: sizes and codes

Size	Small	Medium	Large	X-Large
Code	GL6631	GL6632	GL6633	GL6635

Food Contact



GL663 VitriLe gloves have been tested in accordance with EN1186. They are suitable for use with all food categories, except fatty foods, in situations of short term, repeat contact.

Done at Enfield, 27/01/14

A handwritten signature in black ink, appearing to read 'Bernard Garvey'.

Bernard Garvey
Technology Director