

EC Declaration of Product Conformity

The manufacturer established in the community;

**BM Polyco Ltd
Crown Road
Enfield
EN1 1TX
United Kingdom**

Certified to ISO 9001 & ISO 13485

declares that the new Medical Device and PPE described hereafter



BLUE NITRILE GL900

Non-sterile, powder-free, nitrile disposable glove

is in conformity with the provisions of Council Directive 93/42/EEC and with National Standards transposing harmonised standards EN455-1, EN455-2, EN455-3, EN455-4 and is self-certified as a Class 1 non-sterile medical device

Freedom from holes: AQL 1.5
Physical properties: 6N minimum

is in conformity with the provisions of Council Directive 89/686/EEC and where such is the case, with National Standards transposing harmonised standards EN374-1:2003, EN374-2:2003, EN374-3:2003 and EN420:2003.

is identical to the PPE which is the subject of EC certificate of conformity no. 7050 issued by SATRA Technology Centre, Wyndham Way, Telford Way, Kettering, Northamptonshire, NN16 8SD, United Kingdom (Notified Body No. 0321).

conforms to the provisions of Council Directive 89/686/EEC Article 11B under the supervision of the Notified Body, LRQA UK, 1 Trinity Park, Bickenhill Lane, Birmingham, B37 7ES, UK (Notified Body No. 0088).


**MICRO ORGANISM
HAZARDS
EN374-2**



**CHEMICAL
HAZARDS
EN374-3**



CE 0088



Complex Design

User Information

This product does not contain natural rubber latex.
Glove Care: Store below 40°C away from direct sunlight
You are advised to retain this packaging for reference.
Contains accelerators which may cause allergic reactions.

Range Available: sizes and codes

Size	Small	Medium	Large	X-Large
Code	GL9001	GL9002	GL9003	GL9005

Food Contact



GL900 have been tested in accordance with EN1186. They are suitable for use with all food categories in situations of short term, repeat contact.

Done at Enfield, 02/06/15



Bernard Garvey
Technology Director